

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/10/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:19 |
| INTERNAL STANDARD | VERIFIED | 01:19 |
| REFERENCE STANDARD TEMP | 34.00c | |
| REFERENCE STANDARD | 0.09 | 01:19 |
| BLANK TEST | 0.00 | 01:20 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 01:20 |
| BLANK TEST | 0.00 | 01:21 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 01:21 |
| BLANK TEST | 0.00 | 01:22 |

[Handwritten signature]
#1870

