

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/17/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:16
INTERNAL STANDARD	VERIFIED	22:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:16
BLANK TEST	0.00	22:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:18
BLANK TEST	0.00	22:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:19
BLANK TEST	0.00	22:20

[Handwritten signature]
#820

