

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/02/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:36
INTERNAL STANDARD	VERIFIED	23:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:37
BLANK TEST	0.00	23:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:39
BLANK TEST	0.00	23:40

