

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/14/2025

OPERATOR NAME:
ANTHONY E CARDENAS
Permit #: 50642
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:53
INTERNAL STANDARD	VERIFIED	21:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:53
BLANK TEST	0.00	21:54
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:54
BLANK TEST	0.00	21:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:56
BLANK TEST	0.00	21:56



A. Cardenas #184