

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/21/2025

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:53
INTERNAL STANDARD	VERIFIED	09:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:53
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:54
BLANK TEST	0.00	09:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:55
BLANK TEST	0.00	09:56



P. Bateholts 07/14