

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 03/10/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:11
INTERNAL STANDARD	VERIFIED	20:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:12
BLANK TEST	0.00	20:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:13
BLANK TEST	0.00	20:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:14
BLANK TEST	0.00	20:15

#820

