

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 03/14/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:32
INTERNAL STANDARD	VERIFIED	23:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:32
BLANK TEST	0.00	23:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	23:33
BLANK TEST	0.00	23:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:35
BLANK TEST	0.00	23:36

