

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 04/04/2025

OPERATOR NAME:  
J T COON  
Permit #: 36043  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:32
INTERNAL STANDARD	VERIFIED	10:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:33
BLANK TEST	0.00	10:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:35
BLANK TEST	0.00	10:36

*P.O. J.T. Coon 341*

