

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 04/18/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:31
INTERNAL STANDARD	VERIFIED	08:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:31
BLANK TEST	0.00	08:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:33
BLANK TEST	0.00	08:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:34
BLANK TEST	0.00	08:35

