

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 04/26/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:03
INTERNAL STANDARD	VERIFIED	22:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:04
BLANK TEST	0.00	22:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:06
BLANK TEST	0.00	22:07

#800

