

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 05/02/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:14
INTERNAL STANDARD	VERIFIED	09:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:15
BLANK TEST	0.00	09:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:16
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:17
BLANK TEST	0.00	09:18

*PO Alton*      *Permit # 44983*

