

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/06/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:20
INTERNAL STANDARD	VERIFIED	02:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:21
BLANK TEST	0.00	02:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:22
BLANK TEST	0.00	02:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24

