

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/24/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:08
INTERNAL STANDARD	VERIFIED	20:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:09
BLANK TEST	0.00	20:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:10
BLANK TEST	0.00	20:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:11
BLANK TEST	0.00	20:12

