

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 05/30/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:48
INTERNAL STANDARD	VERIFIED	00:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:48
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:49
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51

[Handwritten signature]
#820

