

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 06/01/2025

OPERATOR NAME:  
JEREMY M CRISPELL  
Permit #: 51155  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:42
INTERNAL STANDARD	VERIFIED	10:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:43
BLANK TEST	0.00	10:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:44
BLANK TEST	0.00	10:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:45
BLANK TEST	0.00	10:46

