

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 06/16/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:47
INTERNAL STANDARD	VERIFIED	23:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:47
BLANK TEST	0.00	23:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:49
BLANK TEST	0.00	23:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:50
BLANK TEST	0.00	23:51

