

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/20/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:54
INTERNAL STANDARD	VERIFIED	19:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:55
BLANK TEST	0.00	19:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:56
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:57
BLANK TEST	0.00	19:58

J Riehl
#1870

