

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 06/28/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:24
INTERNAL STANDARD	VERIFIED	16:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:24
BLANK TEST	0.00	16:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	16:25
BLANK TEST	0.00	16:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:26
BLANK TEST	0.00	16:27

PO Steven T Bunio 789

