

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/13/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:25
INTERNAL STANDARD	VERIFIED	12:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:25
BLANK TEST	0.00	12:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:27
BLANK TEST	0.00	12:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:28
BLANK TEST	0.00	12:29

