

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/18/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:42
INTERNAL STANDARD	VERIFIED	23:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:42
BLANK TEST	0.00	23:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:44
BLANK TEST	0.00	23:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	23:45
BLANK TEST	0.00	23:46

