

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 07/31/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:01
INTERNAL STANDARD	VERIFIED	16:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:02
BLANK TEST	0.00	16:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	16:03
BLANK TEST	0.00	16:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:04
BLANK TEST	0.00	16:05

#820

