

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/15/2025

OPERATOR NAME:
JESSE P SUMMO
Permit #: 47506
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:07
INTERNAL STANDARD	VERIFIED	20:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:07
BLANK TEST	0.00	20:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:09
BLANK TEST	0.00	20:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:10
BLANK TEST	0.00	20:11



JS 174