

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 08/29/2025

OPERATOR NAME:
JESSE P SUMMO
Permit #: 47506
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:34
INTERNAL STANDARD	VERIFIED	21:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:35
BLANK TEST	0.00	21:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:37
BLANK TEST	0.00	21:38

JS 174

