

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/03/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:20
INTERNAL STANDARD	VERIFIED	09:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:20
BLANK TEST	0.00	09:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:21
BLANK TEST	0.00	09:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:23
BLANK TEST	0.00	09:23

*PT Steven Bunio # 44983*

