

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 10/11/2025

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:22
INTERNAL STANDARD	VERIFIED	02:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:23
BLANK TEST	0.00	02:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26

*[Handwritten signature]*  
#820

