

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 10/24/2025

OPERATOR NAME:
MICHAEL A BAKER
Permit #: 49901
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:42
BLANK TEST	0.00	13:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:43
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:45
BLANK TEST	0.00	13:45

Michael A Baker 148

