

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 11/01/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:09
INTERNAL STANDARD	VERIFIED	12:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:10
BLANK TEST	0.00	12:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:11
BLANK TEST	0.00	12:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:12
BLANK TEST	0.00	12:13

*PO Steve Bunio 44983*

