

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 11/08/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:52
INTERNAL STANDARD	VERIFIED	00:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:54
BLANK TEST	0.00	00:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:55
BLANK TEST	0.00	00:56

[Handwritten signature]
#820

