

SOLUTION CHANGE



WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/02/2025

OPERATOR NAME:
ETHAN P OSTAPCZUK
Permit #: 50075
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:02
INTERNAL STANDARD	VERIFIED	21:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:03
BLANK TEST	0.00	21:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:04
BLANK TEST	0.00	21:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:05
BLANK TEST	0.00	21:06



Ethanol 152