

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 12/13/2025

OPERATOR NAME:  
ROBERT A ROSE  
Permit #: 51175  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:04
INTERNAL STANDARD	VERIFIED	03:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	03:05
BLANK TEST	0.00	03:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:07
BLANK TEST	0.00	03:08

*RA Rose #897*

