

SUPERVISOR TEST



WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 12/20/2025

OPERATOR NAME:  
JESSE P SUMMO  
Permit #: 47506  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:43
INTERNAL STANDARD	VERIFIED	00:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:45
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:47



*JS* 174