

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 12/26/2025

OPERATOR NAME:
ETHAN P OSTAPCZUK
Permit #: 50075
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:31
INTERNAL STANDARD	VERIFIED	00:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:34
BLANK TEST	0.00	00:35



Ethan P Ostapczuk 152