

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/02/2025

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:26
INTERNAL STANDARD	VERIFIED	07:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:28
BLANK TEST	0.00	07:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:29
BLANK TEST	0.00	07:30



John D Bateholts #714