

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 01/10/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:33
INTERNAL STANDARD	VERIFIED	00:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:33
BLANK TEST	0.00	00:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:35
BLANK TEST	0.00	00:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:36
BLANK TEST	0.00	00:37

[Handwritten signature]
#820

