

SOLUTION CHANGE

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 02/01/2025

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24110

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:34
INTERNAL STANDARD	VERIFIED	21:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:35
BLANK TEST	0.00	21:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:36
BLANK TEST	0.00	21:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:37
BLANK TEST	0.00	21:38

