

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 02/07/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:12
INTERNAL STANDARD	VERIFIED	10:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:14
BLANK TEST	0.00	10:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:15
BLANK TEST	0.00	10:16

PT Steve Bunio #39

