

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/10/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:37
INTERNAL STANDARD	VERIFIED	22:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	22:38
BLANK TEST	0.00	22:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:39
BLANK TEST	0.00	22:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:40
BLANK TEST	0.00	22:41

[Handwritten signature]
#820

