

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 03/15/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:08
INTERNAL STANDARD	VERIFIED	06:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:08
BLANK TEST	0.00	06:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:10
BLANK TEST	0.00	06:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:11
BLANK TEST	0.00	06:12

