

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/01/2025

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:45
INTERNAL STANDARD	VERIFIED	08:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:47
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:48
BLANK TEST	0.00	08:49



P. Bateholts 0714