

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/12/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:01
INTERNAL STANDARD	VERIFIED	02:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:01
BLANK TEST	0.00	02:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:04

#880

