

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 04/18/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:46
INTERNAL STANDARD	VERIFIED	18:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:47
BLANK TEST	0.00	18:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:48
BLANK TEST	0.00	18:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:49
BLANK TEST	0.00	18:50

