

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 04/27/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:58
INTERNAL STANDARD	VERIFIED	18:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	18:59
BLANK TEST	0.00	19:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:00
BLANK TEST	0.00	19:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:01
BLANK TEST	0.00	19:02

