

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/02/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:19
INTERNAL STANDARD	VERIFIED	07:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:21
BLANK TEST	0.00	07:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:22
BLANK TEST	0.00	07:23

Handwritten signature and permit number 44983

