

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/12/2025

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:48
INTERNAL STANDARD	VERIFIED	05:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:49
BLANK TEST	0.00	05:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:50
BLANK TEST	0.00	05:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:51
BLANK TEST	0.00	05:52

