

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 05/16/2025

OPERATOR NAME:  
MICHAEL A CAMPBELL  
Permit #: 42422  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:53
INTERNAL STANDARD	VERIFIED	05:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:54
BLANK TEST	0.00	05:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:55
BLANK TEST	0.00	05:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:56
BLANK TEST	0.00	05:57

