

SUPERVISOR TEST

 WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 05/24/2025

OPERATOR NAME:
 MICHAEL A CAMPBELL
 Permit #: 42422
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:16
INTERNAL STANDARD	VERIFIED	18:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:17
BLANK TEST	0.00	18:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:18
BLANK TEST	0.00	18:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:19
BLANK TEST	0.00	18:20

