

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/30/2025

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:58
INTERNAL STANDARD	VERIFIED	01:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:58
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:00
BLANK TEST	0.00	02:01

[Handwritten signature]
#880

