

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 06/01/2025

OPERATOR NAME:
JEREMY M CRISPELL
Permit #: 51155
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:58
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:01
BLANK TEST	0.00	10:02



Handwritten signature and date: 06/01/25