

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 06/09/2025

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:42
INTERNAL STANDARD	VERIFIED	11:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:43
BLANK TEST	0.00	11:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:44
BLANK TEST	0.00	11:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:45
BLANK TEST	0.00	11:46

PO Steve Bunio #759

