

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 06/28/2025

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 25510

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:42
INTERNAL STANDARD	VERIFIED	18:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:43
BLANK TEST	0.00	18:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:44
BLANK TEST	0.00	18:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:45
BLANK TEST	0.00	18:46

